
NOVEMBER 2023

Product Bulletin

NGC8200/PGC1000 ASTM D7164-21 Compliance


ABB has reviewed the ASTM D7164-21 and can confirm compliance with the standard, exclusively as it applies to the analyzer's function.

Please note, according to section 8 of ASTM D7164-21, the correct installation of the analyzer and sample system is the responsibility of the end user, rather than the analyzer manufacturer. Additionally, the performance tests in section 9 and subsequent reporting are the responsibility of the end-user. The end user should use appropriate company Standard Operating Procedures (SOP) for integrating data from the NGC into their reports, as this functionality is not within the ABB NGC software.

Respectfully submitted,



Damian C. Huff
Product Line Manager – GC Products



Steve Lakey
Technical Product Manager – GC Products